

***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst

Article number: FD51600

Batch number: 41XI18GA00 Quantity: 9 480

Manufacturing date: 19 September 2025 Expiry date: 2027-08

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA022	

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 13.10.2025	Fresenius Kabi AG
Name: Tykva František	Plant:
Signature: 	Fresenius Kabi Horátev CZ s.r.o.
	Hořátev 104, 289 13
	Czech Republic