



**FRESENIUS
KABI**

***** CERTIFICATE OF ANALYSIS *****

Article: CompoStop Flex2F PLT processing and leuko red.syst

Article number: FD51600

Batch number: 41XI30GA00

Quantity: 5 880

Manufacturing date: 29 September 2025

Expiry date: 2027-08

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA019	

<u>Batch released by Quality Assurance.</u> Date: 17.10.2025 Name: Tykva František Signature: 	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi Horatev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
---	--

Form.no. 17-4269/2