

**** **CERTIFICATE OF ANALYSIS** ****

Article: CompoFlex® 1F Empty - PD-V
Article number: T2131
Batch number: 85XG17CA00 Quantity: 8321
Manufacturing date: 2025.07 Expiry date: 2028.06

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 05 Aug 2025 Name: Xu Xiaoying Signature: Xu Xiaoying	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi (Guangzhou) Co., Ltd.
---	---